

Nodularin Report*Project: Central Davis Sewer District*

Submitted to: Leland Myers
Organization: Central Davis Sewer District
Email: ljmyers@cdsewer.org
Sample Receipt Date: 9 November 17
Sample Condition: 16.2 °C
Report #: 171109 – Central Davis Sewer District
Date Prepared: 17 November 17
Prepared by: Mark Aubel

<u>Sample Identification</u>	<u>Description/Site</u>	<u>Sample Collection Date</u>
FB1	Great Salt Lake	8 November 17
FB4	Great Salt Lake	8 November 17

Analytes: Nodularin (NOD)

Sample Preparation***Water Sample Ultrasonication***

Samples were received and immediately frozen for later preparation. After thawing, the samples were inverted for 60 seconds to mix and sonicated to lyse cells and release of toxins.

Solid Phase Extraction (SPE)

Preconditioned Strata X Polymeric SPE (200 mg) columns were loaded with 2.5 mL of sample, rinsed with 5% MeOH and eluted with 90% acetonitrile. Elutions were blown to dryness (N₂ at 60°C) and reconstituted in Deionized water (0.5 mL, providing a 5x preconcentration).

Quality Control

Table 1: LFSM/LFSMD QC sample prepared for analysis (unless otherwise noted)

Analyte	Concentration (ng/mL)	Sample ID(s)	Return
NOD	0.4	FB1	114%

Additional Quality Control/Quality Assurance checks included method blanks and a LFB.

Analytical Techniques

NOD

The method described in Foss and Aibel (2015) was modified to accommodate only nodularin. A Certified Reference Standard of NOD (1.0 ng/mL) was used to calibrate the method. Table 2 below shows the transition monitored. A MDL was determined through standard addition (LFSM).

Table 2

Analyte	Precursor Ion (<i>m/z</i>)	Fragment Ions (<i>m/z</i>)
NOD	[M+H] ⁺ 825.5	599, 674, 776, 781

Summary of Results

Sample ID	NOD (ng/mL)
FB1	0.16
FB4	0.09
MDL (ng/ mL)	0.05
Analyst Initials	MA
Date Analyzed	11/17/17

Abbreviations:

MDL	Method Detection Limit
MDL	Method Quantification Limit
ND	Not Detected above the MDL
Blank	Regent Water free from interferences
LFB	Lab Fortified Blank
LFSM	Lab Fortified Sample Matrix
LFSMD	Lab Fortified Sample Matrix Duplicate
LD	Lab Duplicate

Submitted by:



Mark T. Aubel, Ph.D.

Date:

November 17, 2017

This report shall not be reproduced except in full without written approval of the laboratory